



The Bellota Sonohysterography Catheter with Cervical Cone is available in a 5(f) size and is 296mm in length.

The catheter is to be used for sonohysterography procedures that aid in the analysis of uterine pathology. A placement sheath eliminates the need for a tenaculum. This catheter is designed for single patient use.

PRODUCT NO. UIC 5FA
UIC 5FAX

Not made with natural rubber latex

PACKAGING: 25 sterile catheters per box

CAUTION

Federal law restricts this device to sale by or on the order of a physician.

Do not reuse for avoiding user may be infected by the microorganism.

INDICATIONS

The Bellota Sonohysterography Catheter with Cervical Cone is used to access the uterine cavity for saline infusion Sonohysterography procedure

CONTRAINDICATIONS

The catheter type procedures are contraindicated in pregnancies, suspected pregnancies, active pelvic infections, or recent pelvic infections, severe uterine bleeding, gynecologic malignancies. This device should not be used for intrafallopian or intratubal procedure or for in-vitro fertilization.

PROCEDURE

- Remove the catheter from the sterile pouch.
- Attach the saline to the open luer and allow the saline to flow through the catheter removing air and flushing the saline channel.

Visit our website:

www.bioteque.com

Tel: 1-800-889-9008

Fax: 1-408-526-9399



Wellkang Ltd
Suite B, 29 Harley Street
LONDON, W1G 9QR, U.K.

- Adjust the protective sheath so that just 1/2 to 1cm of the catheter protrudes beyond the sheath.
- Expose the cervix using a speculum and clean it with an antiseptic solution.
- The single-lumen tube is featured with centimeter depth markings along its distal segment from 6-10cm to set the depth of insertion. It is helpful to place the catheter properly.
- Carefully advance the 5f catheter through the cervical canal and past the internal cervical OS, so that the catheter is within the uterine cavity while the catheter advanced and placing the acorn to avoid any fluid reflux.
- Attach saline solution to open syringe and introduce through catheter slowly.
- Remove the speculum and insert the transvaginal transducer into its proper position adjacent to the catheter.
- Now activate the ultrasound system and inject saline as required by the procedure. Usually, 3cc to 20cc is sufficient saline for a proper diagnosis.
- Perform Sonohysterography in the usual way.
- Endometrial distention usually can be maintained throughout the procedure. If uterine distention is not visualized with ultrasound while saline is injected into the catheter or leaking is observed from the vagina, verify that the catheter is positioned correctly in the uterine cavity before continuing.
- Upon completion of Sonohysterography, remove the catheter carefully from uterine cavity.

STORAGE: Store at room temperature.



DO NOT REUSE



USE BY



BATCH CODE



STERILE-ETHYLENE OXIDE

Manufacturer:



Panpac Medical Corporation
6F-2, No. 202, Sec. 3, Ta-Tong
Road, Shi-Chih District,
New Taipei City, Taiwan